


VIDEO CYSTOSCOPE

S2+ Series




Model: OCY-S2+



LED At Tip



CMOS sensor





Laser compatible




Single Step,
Water Proof
Connector

Features:

- LED-AT-TIP Technology for longer scope life.
- CMOS Sensor provides unmatched image quality to explore smallest details.
- Ergonomic, slim and flexible for smooth insertion.
- Easy manoeuvrability.
- Laser compatible.
- Can be used for examination / procedure in Kidneys
- Bladder and a wide range of surgical procedures.
- 120° field of view captures both ureteral openings.
- Flexible angulation (210° Up and 130° Down) facilitates smooth orientation in the urinary bladder and easier observation of the bladder neck.



Specifications:

| | | |
|-----------------------------------|--|---|
| Model | OCY-S2+ | |
| Optical System | Field of view | 120° |
| | Direction of view | Forward View |
| | Depth of field | 3-60mm |
| Insertion Tube | Distal end outer diameter | 2.80 mm Bullet Shape |
| | Insertion tube outer diameter | 4.9 mm |
| | Distal end enlarged | 1. Illumination LEDs (W) |
| | | 2. Image Sensor |
| | | 3. Instrument Channel Outlet |
| Instrument Channel | Working Length | 380 mm |
| | Channel inner diameter | 2 mm |
| | Minimum Visible Distance | 3 mm from the distal end |
| | Direction from which endotherapy accessories enter and exit the endoscopic image |  |
| Bending Section | Angulation range | UP 210°, DOWN 130° |
| Total Length | | 675 mm |
| Laser Compatibility | | Nd: YAG, 810nm Diode |
| Compatible Video Processor | | OVP-S2+ |

*Specifications are subject to change.



Manufactured by:

MITRA MEDICAL SERVICES LLP

Plot No.: 21-B, Sector-1, Parwanoo, Dist.- Solan, Himachal Pradesh - 173 220. INDIA

Toll Free No.: 1800 1020 307

E-Mail: ottomed@ottomed.com, Website: www.ottomed.com

Distributed by:

MITRA MEDICAL SERVICES LLP

B-226, Okhla Industrial Area, Phase-1, New Delhi - 110 020. INDIA

Toll Free No.: 1800 1020 307

E-Mail: ottomed@ottomed.com, Website: www.ottomed.com

Printed - 20 SEPTEMBER 2021